

LeviCell EOS

TABLE OF CONTENTS

Table of Contents..... 1

Site Requirements..... 2

Pre-Installation Questionnaire..... 3

LeviCell EOS Specifications..... 4

Legal..... 5

Contact Information..... 5

SITE REQUIREMENTS

Loading dock and access

- The LeviCell EOS arrives with the instrument module in a cardboard box on a pallet, with an overall dimension of 1120 mm H x 620 mm W x 950 mm L (44.1" H x 24.4" W x 37.4" L) and a gross weight of approximately 100 kg (220 pounds). This includes two containers:
 - Container 1: LeviCell EOS Module, with the dimensions of 749 mm H x 616 mm W x 946 mm L (29.5" H x 24.25" W x 37.25" L) and a gross weight of approximately 65 kg (143 pounds).
 - Container 2: The Control PC and its accessories, with the dimensions of 371 mm H x 616 mm W x 946 mm L (14.6" H x 24.25" W x 37.25" L) and a gross weight of approximately 35 kg (77 pounds).
- A hand-operated pallet truck is required. Please contact LevitasBio if you require any different delivery arrangements.
- The instrument crate will need to be moved to the lab before the day of installation. Verify that all doors are wide enough to admit the crate on the pallet jack.
 - The shortest width of the instrument crate, by itself, will fit through a standard-width door.
- For installation, the instrument weighs about 45.4 kg (100 pounds); this requires a 2-person lift from the shipping container onto the bench. This step will be performed by LevitasBio personnel.

Bench space: available bench space should be at least:

- 660 mm (26") deep, 1200 mm (48") wide.
- A 1520 mm (60") wide bench is suggested, with 508 mm (20") vertical space available above the bench.
- Minimum 3" clearance behind the instrument for venting and access to the main power button and enough clearance on the sides to reach the main power button on the rear of the instrument.

Power: Access to 2 standard power outlets (or an outlet strip, although a direct connection is recommended).

- All standard mains input voltages are accepted (100 - 240 VAC, 50-60 Hz)
- Power drawn is low, so no special circuits are required.
- Grounded outlets are required (3-pin, or CEE 7/3 style)

Installation and Usage:

- The system will be installed by LevitasBio personnel. The system box should not be opened prior to installation.
- It is recommended that the unit be placed such that the back power switch (right side of the instrument when viewed from the front) is easily accessible. This controls the main power of the instrument.
- Do not obstruct access to the mains power supply cable leading into the instrument or the end of the cord at its inlet. It should be removable if power needs to be completely removed for servicing.

Front/Side View



Rear/Side View



- A soft power button is present on the front panel of the instrument.
- In the case that immediate power shutdown is required, use the rear mains power switch or remove the mains plug.
- The system covers should always remain in place to avoid mechanical hazards to the user.
- When the system is in use, the user must avoid the clamping mechanism WHEN it engages the cartridge to avoid pinch hazards. Handling the cartridge by its grip will avoid the hazard.

PRE-INSTALLATION QUESTIONNAIRE

Installation item	Site status
Loading dock access or lift-gate required?	
Any special delivery requirements?	
Any special requests?	

LEVICELL EOS SPECIFICATIONS

Specification	Value
Number of sample inputs	4
Number of output fractions	8
Levitation magnets	Rare earth permanent magnets
Separation flow rate	300 µL/min
Imaging modes	Brightfield (transmitted illumination) Two fluorescence channels: Excitation 470 nm, Emission 501-544 nm (e.g., Calcein, FITC) Excitation 560 nm, Emission 601-666 nm (e.g., PI)
Imaging resolution	Approximately 2 microns
Operational	
Input voltage	100 - 240 VAC nominal, universal 50 - 60 Hz (90 - 264 VAC maximum operating range) Standard wall receptacle (over-voltage category II)
AC supply current drawn	4A
Main enclosure dimensions	440 W x 630 D x 460 H mm (17.3" W x 24.8" D x 18.1" H)
Instrument weight	45.4 kg (100 pounds)
Control PC Operating System	Windows 10 Professional, 64 bit
Ingress protection rating	Not rated (no protection claimed)
Environmental	
Operating ambient temperature	19°C – 25°C
Operating relative humidity	20% RH – 80% RH ambient, non-condensing
Pollution degree of the intended environment	Pollution Degree 2 (typical indoor laboratory environment)
Altitude	Sea level to 2,000m (6,562 feet)
For indoor use only	Not designed for outdoor use. Not designed for use in wet locations.
Shipping environment	5 °C to 50 °C, 5% - 99% RH, non-condensing

LEGAL

© 2023 LevitasBio, Inc. All rights reserved. “Levitas”, “LevitasBio”, and “LeviCell” are trademarks of LevitasBio, Inc. All other trademarks are the property of their respective owners. All products and services described herein are for Research Use Only. Not for use in diagnostic procedures.

The LeviCell® EOS system, is developed by and manufactured for LevitasBio, Inc. This device and/or methods of using this device are covered by one or more issued patents and pending patent applications owned or licensed by LevitasBio, Inc., including US Patent Numbers 9,517,474, 9,873,126, and 10,928,404. LevitasBio products may not be transferred to third parties, resold, modified for resale, or used to manufacture commercial products without prior written approval of LevitasBio, Inc.

Your use of this product is subject to Sales Terms and Conditions at <https://levitasbio.com>.

This manual provides information about the use of the LeviCell EOS system. LevitasBio shall not be liable for errors contained herein, and shall not be liable for any direct, incidental or consequential damages in connection with or arising from the furnishing, performance, or use of this document or the device, including the loss of any cell samples processed in the device. .

The information contained in this document is subject to change without notice.

CONTACT INFORMATION

Technical Support Number	+1-408-663-4260
Technical Support Email	support@levitasbio.com
Company Website	http://levitasbio.com/
Company Address	LevitasBio 1505 Adams Drive, Suite D Menlo Park, CA 94025, USA